

Purchase Order:

Standard Meter Lab, Inc.

Cert #: 251760



Certificate of Calibration



Customer:

Contact:

#US4461c

Address:

Instrument Identification

System ID: 1057503

Serial #: 00102023000018

Tool #: N/A

Property #: N/A

Manufacturer: PROTRAN

Model #: PROHOT STICK+ (PT-0010)

Range: 50 TO 1500 VOLTS DC

Description: VOLTAGE AWARENESS DEVICE

Test Results

Serviced Performed: Calibration

Service Technician: Keith S Lam

Cal Date: 04/23/2026

Cal. Due Date: 04/23/2028

Location of Cal: In-house

Laboratory: Standard Meter Lab

Address: 236 Rickenbacker Cir, Livermore, CA 94551

As Found Result: In Tolerance

As Left Result: In Tolerance

Environmental Conditions: 68.0 °F / 40.0% RH

Instruction Used: SPEC_PRO-HOTSTICK

Technical Remarks

Condition

Received in fair condition.

Analysis

Verified accuracy in accordance with the listed calibration instructions.

Calibration Standards

I.D.	Manufacturer	Model Number	Description	Cal. Due Date	NIST #
1000154	FLUKE ELECTRONICS	5522A/GHZ (ACCREDITED)	CALIBRATOR, MULTI FUNCTION	2/5/2027	248375

Calibrations are performed using standards traceable to NIST. Our calibration system complies with ISO/IEC 17025:2017. This information applies only to the instrument identified above and may not be reproduced, except in full, without prior written consent. Statements of conformity are reported using the Simple Acceptance Decision Rule (shared risk) in accordance with ILAC G8:09/2019. For this rule the following criteria is applied: Out of Tolerance(Yes)-Measured Value exceeds given specifications. Out of Tolerance(No)-Measured Value is within given specifications. The presence of the Intertek logo designates our quality management system is certified to ISO 9001:2015. The reported expanded measurement uncertainty is stated as the standard measurement uncertainty multiplied by the coverage factor K such that coverage probability corresponds to approximately 95%. There is no implied warranty that the instrument will maintain its specified tolerances during the calibration interval due to possible drift, environment or other factors beyond our control.

Approval Person: Alejandra Reyes Quality Assurance

Signature: Alejandra Reyes

Date: 04/23/2026